

## UK Declaration of Conformity

The Non-Automatic Weighing Instrument

III

|                                  |                             |
|----------------------------------|-----------------------------|
| Manufacturer                     | Charder Electronic Co., Ltd |
| Model                            | M-545                       |
| Type Examination Certificate No. | T10991-UK                   |

### The Metrological Aspects of Non-Automatic Weighing Instruments

|                          |                             |
|--------------------------|-----------------------------|
| EN45501: 2015 (module D) | Approved Body Number – 8506 |
| EN45501: 2015 (module B) | Approved Body Number – 8506 |

The non-automatic weighing instrument corresponds to the production model described in the Type Examination Certificate and requirements of the following UK Regulations:

|                   |   |
|-------------------|---|
| S.I. 2016 No.1152 | Non-Automatic Weighing Instruments Regulations 2016 |
|-------------------|---|

The applicable designated standards are:

|   |   |
|---|---|
| EN45501:2015                                      | The Metrological Aspects of Non-Automatic Weighing Machines   |
| EN 60601-1:2006+A11:2011+A1:2013+A12:2014+A2:2021 | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance  |
| EN 60601-1-2:2015+A1:2021                         | Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests |

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Date: 2024.09.04

Signature: *Victor Lai*

Name: Victor Lai  
Position: Measuring Management Rep.  
Place: Taichung, Taiwan

Manufacturer: Charder Electronic Co., Ltd.

Address: NO.103, Guozhong Rd., Dali Dist., Taichung City 412, Taiwan