

UK Declaration of Conformity

The Non-Automatic Weighing Instrument

III

| Manufacturer | Charder Electronic Co., Ltd |
|----------------------------------|-----------------------------|
| Model | M-900 |
| Type Examination Certificate No. | T7616-UK |

The Metrological Aspects of Non-Automatic Weighing Instruments

| EN45501: 2015 (module D) | Approved Body Number – 8506 |
|--------------------------|-----------------------------|
| EN45501: 2015 (module B) | Approved Body Number – 8506 |

The non-automatic weighing instrument corresponds to the production model described in the Type Examination Certificate and requirements of the following UK Regulations:

The applicable designated standards are:

| EN45501:2015 | The Metrological Aspects of Non-Automatic Weighing Machines |
|---------------------------|---|
| EN 60601-1:2006+A11:2011+ | Medical electrical equipment - Part 1: General requirements for basic |
| A1:2013+A12:2014+A2:2021 | safety and essential performance |
| EN 60601-1-2:2015+A1:2021 | Medical electrical equipment - Part 1-2: General requirements for basic |
| | safety and essential performance - Collateral Standard: Electromagnetic |
| | disturbances - Requirements and tests |

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Date: 2024.09.04 Signature: Victor Lai

Name: Victor Lai

Position: Measuring Management Rep.

Place: Taichung, Taiwan

Manufacturer: Charder Electronic Co., Ltd.

Address: NO.103, Guozhong Rd., Dali Dist., Taichung City 412, Taiwan