




Stand-on Floor Scale

USER MANUAL

MS4203

Stand-on Floor Scale

















Please keep the instruction manual at hand and follow instruction for use.


CONTENTS

| | |
|---|-----------|
| I. Explanation of Graphic Symbols on Label/Packaging.. | 3 |
| II. Copyright Notice..... | 5 |
| Taichung City, 41262 Taiwan | 5 |
| III. Safety Notes | 6 |
| A. General Information..... | 6 |
| B. EMC Guidance and Manufacturer's Declaration | 9 |
| IV. Installation | 13 |
| A. Inserting Batteries..... | 13 |
| B. Using Adapter | 14 |
| V. Indicator | 15 |
| VI. Using Device | 16 |
| A. Basic Operation | 16 |
| B. Zero..... | 17 |
| VII. Troubleshooting | 18 |
| VIII. Product Specifications | 22 |
| A. Device Information | 22 |
| B. Power Adapter Standards | 23 |
| IX. Declaration of Conformity..... | 24 |

I. Explanation of Graphic Symbols on Label/Packaging

| Text/Symbol | Meaning |
|---|---|
|  | Caution, consult accompanying documents before use |
|  | Separate collection for waste of electrical and electronic equipment, in accordance with Directive 2002/96/EC. Do not dispose of device with everyday waste |
|  | Name and address of device manufacturer, and year/country of manufacture |
|  | Carefully read user manual before installation and usage, and follow instructions for use. |
|  | Medical electrical device, Type B applied part |
|  | Medical electrical device, Type BF applied part |
|  | Device catalogue number / model number |
|  | Name and address of authorized representative in the European Union |
|  | Device is a medical device. Text indicates device category type |
|  | Manufacturer's batch or lot number for device |
|  | Device's serial number |
|  | Device's Unique Device Identifier |
|  | Verification Scale Interval. Value expressed in units of mass. Used to classification and verification of an instrument. |
|  | Device conforms to (EU) 2017/745 Regulation on Medical Devices. Four digit number is identifier for medical device Notified Body |

Device complies with EC directives (verified models only)

 **M**201222
M: Conformity label in compliance with Directive 2014/31/EU for non-automatic weighing instruments
20: Year in which conformity verification was performed and the CE label was applied. (ex: 16=2016)
0122: Identifier for metrology Notified Body



Device is a Class III scale in compliance with Directive 2014/31/EU (verified models only)



Name and address of entity importing device (if applicable)



Name and address of entity responsible for translating Information For Use (if applicable)

CON.


Event counter confirming how many times device has been calibrated (if applicable)



Device conforms to Taiwan National Communications Commission(NCC) approval



Device conforms to U.S. Federal Communications Commission regulations

 **M**208506

Device complies with UK non-automatic weighing instruments regulations 2016 (verified models only)

M: Conformity label in compliance with Non-automatic Weighing instruments Regulations 2016
20: Year in which conformity verification was performed and the UKCA label was applied. (ex: 20=2020)
8506: Identifier for metrology approved body



Device complies with all UK applicable product legislation



Device's polarity of power.

"In case of differences, icon on device itself takes precedence"

II. Copyright Notice

Copyright Notice Charder Electronic Co., Ltd.

No.103, Guozhong Rd., Dali Dist., Taichung City 41262 Taiwan

Tel: +886-4-2406 3766

Fax: +886-4-2406 5612

Website: www.chardermedical.com

E-mail: info_cec@charder.com.tw

Copyright© Charder Electronic Co., Ltd. All rights reserved.
This user manual is protected by international copyright law. All content is licensed, and usage is subject to written authorization from Charder Electronic Co., Ltd. (hereinafter Charder) Charder is not liable for any damage caused by a failure to adhere to requirements stated in this manual. Charder reserves the right to correct misprints in the manual without prior notice, and modify the exterior of the device for quality purposes without customer consent.



Charder Electronic Co., Ltd.
No. 103, Guozhong Rd., Dali Dist.,
Taichung City, 41262 Taiwan

III. Safety Notes

A. General Information

Thank you for choosing this Charder Medical device. It is designed to be easy and straightforward to operate, but if you encounter any problems not addressed in this manual, please contact your local Charder service partner.

Before beginning operation of the device, please read this user manual carefully, and keep it in a safe place for reference. It contains important instructions regarding installation, proper usage, and maintenance.

Intended Purpose

This medical device is designed to be used in accordance with national regulations, to measure weight within specifications, for weight-related usage by professionals.

Clinical Benefit

Measurement results can be used by professionals to diagnose (and monitor) weight-related issues.

Intended medical indications/contraindications

Measurement: patient's body weight. No known contraindications to measurement of body weight.

Intended patient profile

- (a) Age: no restrictions
- (b) Weight: no restrictions within device weight capacity
- (c) Patient Conditions: require measurement of body weight. Able to stand independently without support.

Intended user profile

- (a) At least 20 years old
- (b) Minimum knowledge:
 - To be able to read at a high-school level and understand Arabic numerals (e.g. 1, 2, 3, 4...)
 - Basic hygiene knowledge
 - Trained in device's operation
 - Read the instruction manual
- (c) Language
 - Able to read the language of instruction manual and on-screen

instructions

(d) Qualifications

- No special certifications or qualifications required

Residual Risk Evaluation

(a) All foreseeable risks have been evaluated and considered acceptable. Generally speaking, the most likely risk caused by incorrect usage of the device is less accurate measurement (or inability to use device to acquire measurement), which does not pose imminent physical risk to patient or user.

(b) Benefit-risk ratio is considered acceptable. Stand-on floor scales are an important option for measuring patients. Usage of device is unlikely to result in harm to user or patient.

General Handling

- Ensure all parts are properly locked and tightened before operating the device.
- Measurement accuracy requires the subject's feet, back, and head to be straightly aligned. Please note that height can vary throughout the day
- **CAUTION:** Do not use next to equipment that may cause electromagnetic or other types of interference.

Safety Instructions

Before putting device into use, please read this user manual carefully. It contains important instructions for installation, usage, and maintenance of device.

The manufacturer shall not be liable for damages caused by failure to heed the following instructions:

- The device has an expected service life of 5 years when correctly handled, serviced, and periodically inspected in accordance with manufacturer's instructions.
- Improper installation will render the warranty null and void.
- Observe permissible ambient temperatures for use

Cleaning

- Device surface should be cleaned using alcohol-based wipes.

Maintenance

Please contact your local Charder distributor for regular maintenance and calibration, regular checking of accuracy is recommended; frequency to be determined by level of use and state of device.

Warranty/Liability

- The period of warranty shall be eighteen(18) months, beginning on the date of purchase. Please retain your receipt as proof of purchase.
- No responsibility shall be accepted for damage caused through any of the following reasons: unsuitable or improper storage or use, incorrect installation or commissioning by the owner or third parties, natural wear and tear, changes or modifications, incorrect or negligent handling, chemical, electrochemical, or electrical interference, unless damage is attributable to negligence on the part of Charder.
- This device does not contain any user-maintained parts. All maintenance, technical inspections, and repairs should be conducted by an authorized Charder service partner, using original Charder accessories and spare parts. Charder is not liable for any damages arising from improper maintenance or usage. Dismantlement of the device will void the warranty.

Incident Reporting

Any serious incident that has occurred in relation to the device should be reported to the manufacturer, EU representative (if device is used in EU member state), and competent authority of user/subject's member state.

B. EMC Guidance and Manufacturer's Declaration

| Guidance and manufacturer's declaration-electromagnetic emissions | | |
|--|-------------------|---|
| <p>The product is intended for use in the electromagnetic environment specified below. The customer or the user of the product should assure that it is used in such an environment.</p> | | |
| Emission test | Compliance | Electromagnetic environment-guidance |
| RF emissions CISPR 11 | Group 1 | The product uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. |
| RF emissions CISPR 11 | Class A | The product is suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes. |
| Harmonic emissions IEC 61000-3-2 | Class A | |
| Voltage fluctuations /flicker emissions IEC 61000-3-3 | Compliance | |

Guidance and manufacturer's declaration-electromagnetic immunity


The product is intended for use in the electromagnetic environment specified below. The customer or the user of the product should assure that it is used in such an environment.

| Immunity test | IEC 60601 test level | Compliance level | Electromagnetic environment-guidance |
|---|---|---|--|
| Electrostatic discharge(ESD) IEC 61000-4-2 | <u>± 8 kV contact</u> <u>± 2 kV, ± 4 kV, ± 8 kV, ± 15 kV air</u> | <u>± 8 kV contact</u> <u>± 2 kV, ± 4 kV, ± 8 kV, ± 15 kV air</u> | Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30% |
| Electrical fast transient/burst IEC 61000-4-4 | <u>± 2kV for power supply lines</u> | <u>± 2kV for power supply lines</u> | Mains power quality should be that of a typical commercial or hospital environment. |
| Surge IEC 61000-4-5 | <u>± 1kV line(s) to line(s)</u> <u>± 2kV line(s) to earth</u> | <u>± 1kV line(s) to line(s)</u> <u>± 2kV line(s) to earth</u> | Mains power quality should be that of a typical commercial or hospital environment. |
| Voltage Dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11 | <u>0% UT for 0.5 cycle</u> <u>0% UT for 1 cycle</u> <u>70% UT(30% dip in UT) for 25cycles</u> <u>0% UT for 5 s</u> | <u>0% UT for 0.5 cycle</u> <u>0% UT for 1 cycle</u> <u>70% UT(30% dip in UT) for 25cycles</u> <u>0% UT for 5 s</u> | Mains power quality should be that of a typical commercial or hospital environment. If the user of the product requires continued operation during power mains interruptions, it is recommended that the product be powered from an uninterruptible power supply or a battery. |
| Power frequency (50, 60 Hz) magnetic field IEC 61000-4-8 | <u>30 A/m</u> | <u>30 A/m</u> | The product power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment. |
| NOTE UT is the a.c. mains voltage prior to application of the test level. | | | |

Guidance and manufacturer's declaration-electromagnetic immunity

The product is intended for use in the electromagnetic environment specified below.

The customer or the user of the product should assure that is used in such and environment.

| Immunity test | IEC 60601 test level | Compliance level | Electromagnetic environment-guidance |
|----------------------------------|---|---|---|
| Conducted RF IEC 61000-4-6 | 3 Vrms 150 KHz to 80 MHz <u>6 V in ISM bands between 0,15 MHz and 80 MHz</u> <u>80 % AM at 1 kHz</u> | 3 Vrms 150 KHz to 80 MHz <u>6 V in ISM bands between 0,15 MHz and 80 MHz</u> <u>80 % AM at 1 kHz</u> | Portable and mobile RF communications equipment should be used no closer to any part of the product including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance: $d = 1,2 \sqrt{P}$ $d = 2,3 \sqrt{P}$ 80MHz to 800 MHz $d = 2,3 \sqrt{P}$ 800MHz to 2,7GHz Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres(m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, ^a should be less than the compliance level in each frequency range ^b Interference may occur in the vicinity of equipment marked with the following symbol:  |
| Radiated RF IEC 61000-4-3 | 3 V/m <u>80MHz to 2,7 GHz</u> | 3 V/m <u>80MHz to 2,7 GHz</u> | |

NOTE1 At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

- a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the product is used exceeds the applicable RF compliance level above, the product should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the product.
- b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3V/m.

**Recommended separation distance between
portable and mobile RF communications equipment and the product**

The product is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the product can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the product as recommended below, according to the maximum output power of the communications equipment.

| Rated maximum output power of transmitter W | Separation distance according to frequency of transmitter m | | |
|--|---|--|---|
| | 150 kHz to 80 MHz $d = 1,2\sqrt{P}$ | 80 MHz to 800 MHz $d = 1,2\sqrt{P}$ | 800 MHz to 2,7 GHz $d = 2,3\sqrt{P}$ |
| 0,01 | 0,12 | 0,12 | 0,23 |
| 0,1 | 0,38 | 0,38 | 0,73 |
| 1 | 1,2 | 1,2 | 2,3 |
| 10 | 3,8 | 3,8 | 7,3 |
| 100 | 12 | 12 | 23 |

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres(m) can be estimated using the equation applicable to the frequency of the transmitter, where p is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.


NOTE2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

IV. Installation


Device can be used once batteries are installed (or adapter) is plugged in.

A. Inserting Batteries


1. Locate battery cover on bottom of device



2. Remove battery cover. Insert batteries. Ensure polarity is correct.




3. Insert all batteries. Close cover and turn scale right-side up.



B. Using Adapter

Device can be powered via adapter using port located at rear of device. Plug adapter into device before plugging into mains.



V. Indicator

Indicator and Key Functions



Key Functions

1. On/Zero/Off: Turn device on and off. Zero scale ($\pm 2\%$ of full capacity).
Press and hold to turn off.

Indicator Symbols

2. Zero Indicator: Device is at zero
3. Level Indicator: Determine if device is level
4. Scale Specification: Capacity and graduation of device
5. Hold(optional): Determine if weight lock mode (hold) is active.


VI. Using Device

A. Basic Operation

There are 2 methods to switch on the device:

Auto-on

1. Step onto the device for 2 seconds to turn on device
2. Leave measurement platform



3. The device will automatically perform self-calibration.

Manual switch-on

1. Press **[On/Zero/Off]** key to turn the device on.
2. The device will automatically perform self-calibration.

Note: Do not stand on the scale while the scale is performing self-calibration.

Once "0.00 kg" appears on indicator, device is ready for measurement.

Note: If "0.00 kg" does not display on indicator, press **[On/Zero/Off]** key to zero the device. This function can be used for weight within $\pm 2\%$ of full capacity.

Note: If subject's weight exceeds scale capacity, indicator

will display "Err" prompt due to overload.

B. Zero

The zero function allows the user to clear the residual value of objects from the device's measurement result. Zero can be used when weight of object is \leq at/below 2% of full capacity.

VII. Troubleshooting

Product Defects

Charder's warranty is effective for the original purchaser of this device, subject to the terms and conditions listed in the Warranty Program & Return Policy.

1. If Charder is responsible for a fault or defect present upon receipt of the unit, Charder shall either repair the fault, or supply a replacement unit. Should the repairs or replacement delivery fail, statutory provisions shall be valid. The period of warranty shall be two years, beginning on the date of purchase. Please retain your receipt as proof of purchase.

2. No responsibility shall be accepted for damage caused through any of the following reasons: unsuitable or improper storage or use, incorrect installation or commissioning by the owner or third parties, natural wear and tear, changes or modifications, incorrect or negligent handling, chemical, electrochemical, or electrical interference, unless damage is attributable to negligence on the part of Charder.

If device is not covered under warranty, a service maintenance charge will apply, plus cost of replacement parts.

Before contacting your local Charder distributor for repair service, we recommend considering the following troubleshooting procedures:

Self-inspection

1. Device will not power on

- If battery power is depleted, replace with new batteries
- If batteries are not used, check if the power adapter is plugged into the device properly. Check if power adapter is plugged into mains properly

2. Indicator showing "0000" ZERO SPAN out of range

- Interference due to factors such as RF disturbance or ground vibration. Relocate device to location without interference and try again
- Unstable platform. Relocate device to stable location and try again
- External objects interfering with measurement platform. Clear platform of objects and try again
- Device may not function properly on soft surfaces such as carpets or lawns. Relocate device to location with solid, stable floor
- If the steps above cannot resolve the problem, re-calibration may be required to correct weighing accuracy

Distributor support required

If the following errors occur, we recommend contacting your local Charder distributor for repair or replacement services:

1. Device will not power on





- Faulty on/off key
- Broken or damaged wires causing short circuit or faulty connection
- Safety fuse burnout
- Faulty adapter

2. Indicator damage

Possible hardware defects include: faded LCD screen, blurred text, smeared rainbow screen, incorrect decimal display

- Unable to read data
- Indicator shows "Err.E" after device is switched on
- Keys not responding
- Buzzer malfunction

Error Messages

| Error Message | Reason | Action |
|---|---|---|
|  | Low battery warning Voltage of battery is too low to operate device | Replace batteries, or plug in adapter |
|  | Overload Total load exceeds device's maximum capacity | Reduce weight on measurement platform and try again |
|  | Counting Error Signal from loadcells too low | Error normally caused by faulty loadcell or wiring. Please contact distributor. |
|  | Weight on measurement platform when device is turned on | Remove all objects and restart the scale |

VIII. Product Specifications

A. Device Information


| Model | | MS4203 |
|------------------------------|-------------------|---|
| Weight Measurement | Capacity | 0-150 kg x 50g 150-220 kg x 100g |
| | Accuracy | ±1.5 |
| | LCD Screen | 1.0-inch LCD screen (5 digits) |
| | OIML | Class III |
| Device Dimension | | 325(W) x 310(D) x 65(H) mm |
| Device Weight | | 3.3 kg (Without Battery) |
| Key Functions | | On/Zero/Off |
| Power Supply | | 6 AA batteries / Power adapter |
| Operation Environment | | +5°C~+35°C 15% / 85% RH 700 hPa ~1060 hPa |
| Standard Accessories | | User manual, Power adapter |
| Optional Accessories | | Carrying bag |

B. Power Adapter Standards





Warning

The device is only compatible with the manufacturer's power adapters specified below

| AMP VOLTAGE | DRAWING NO. | CE APPROVED TYPE NO. / MODEL NO. | TYPE | Adapter plug |
|--------------------|--------------------|---|-------------|---|
| 12V 1A | CD-AD-0044 | UES12LCP-120100SPA | US |  90 - degree |
| | | | EU | |
| | | | UK | |
| | | | AU | |

IX. Declaration of Conformity

This product has been manufactured in accordance with the harmonized European standards, following the provisions of the below stated directives:

| | |
|---|---|
|  | (EU) 2017/745 Regulation on Medical Devices |
|  | 2014/31/EU Non-automatic Weighing Instruments Directive (OIML models only) |

RoHS Directive 2011/65/EU and Delegated Directive (EU) 2015/863

Radio Equipment Directive 2014/53/EU

(applicable if wireless module is used)

Part 15 of the Federal Communications Statement Rules

This device may not cause harmful interference.

This device must accept any interference received, including interference that may cause undesired operation.

Please see separate document showing on sticker of device for above markings.

Authorized EU Representative:



Obelis s.a.

Bd Général Wahis, 53
B-1030 Brussels
Belgium

Manufactured by:



Charder Electronic Co., Ltd.
No.103, Guozhong Rd., Dali Dist.,
Taichung City 41262 ,Taiwan

CD-IN-00503 REV004 08/2024